

**Thermo Scientific  
Laboratory  
Temperature  
Control Products**

Manual Part Number U01046  
Rev. 07/07/10

## **ADVANCED Series Thermostats**

**AC 150**

**AC 200**

## **ARCTIC Series Refrigerated Bath Circulators**



**A 5B          A 10**

**A 10B        A 25**

**A 25B        A 28**

**A 28F        A 24B**

**A 40**

## **GLACIER Series Refrigerated Bath Circulators**

**G 50**

## **SAHARA Series Heating Bath Circulators**

**S 3            S 12T        S 14P**

**S 7            S 19T        S 21P**

**S 13**

**S 15**

**S 21**

**S 30**

**S 45**

**S 49**

Visit our Web site at:

<http://www.thermoscientific.com/tc>  
Product Service Information, Applications  
Notes, MSDS Forms, e-mail.

## **Thermo Fisher Scientific**

25 Nimble Hill Road  
Newington, NH 03801  
Tel : (800) 258-0830 or  
(603) 436-9444  
Fax : (603) 436-8411  
[www.thermoscientific.com/tc](http://www.thermoscientific.com/tc)

## **Sales, Service, and Customer Support**

25 Nimble Hill Road  
Newington, NH 03801  
Tel: (800) 258-0830  
Sales: 8:00 am to 5:00 pm  
Service and Support: 8:00 am to 6:00 pm Monday  
through Friday (Eastern Time)  
Fax: (603) 436-8411  
[service.tc.us@thermofisher.com](mailto:service.tc.us@thermofisher.com)

Dieselstrasse 4  
D-76227 Karlsruhe, Germany  
Tel : +49 (0) 721 4094 444  
Fax : +49 (0) 721 4094 300  
[info.tc.de@thermofisher.com](mailto:info.tc.de@thermofisher.com)

### **Statement of Copyright**

Copyright © 2010 Thermo Fisher Scientific. All rights reserved.

This manual is copyrighted by Thermo Fisher Scientific.

Users are forbidden to reproduce, republish, redistribute, or resell any materials from this manual in either machine-readable form or any other form.

# Table of Contents

<b>Preface</b>	<b>i</b>
Compliance .....	i
WEEE .....	i
After-Sale Support .....	ii
Feedback .....	ii
Warranty .....	ii
Unpacking .....	ii
<b>Section 1 Safety</b>	<b>1-1</b>
Safety Warnings .....	1-1
<b>Section 2 General Information</b>	<b>2-1</b>
Description .....	2-1
Thermostat Specifications .....	2-1
Bath/Circulator Specifications .....	2-3
Wetted Materials .....	2-4
<b>Section 3 Installation</b>	<b>3-1</b>
Ambient Conditions .....	3-1
Ventilation .....	3-2
Electrical Requirements .....	3-2
External Circulation .....	3-5
Remote Temperature Sensor .....	3-5
Multifunction Port (AC 200 only) .....	3-5
USB Port (AC 200 only) .....	3-5
Tubing Requirements .....	3-6
Tubing .....	3-7
Fluids .....	3-8
Additional Fluid Precautions .....	3-9
Filling Requirements .....	3-12
Draining .....	3-12
<b>Section 4 Operation</b>	<b>4-1</b>
ADVANCED Thermostat .....	4-1
Setup .....	4-2
Initial Start Up .....	4-2
Daily Start Up .....	4-3
Status Display .....	4-3
Changing the Setpoint .....	4-4
Menu Displays .....	4-5
Menu .....	4-6
Menu Tree .....	4-7

	Editor - Setpoints.....	4-8
	Editor - Ramp Program.....	4-9
	Running a Ramp Program.....	4-11
	Editor - Auto start.....	4-13
	Editor - Timer start.....	4-13
	Settings - Basic settings.....	4-14
	Settings - App. Settings - Limits.....	4-15
	Settings - App. Settings - Fluids Type.....	4-17
	Settings - Password/Reset.....	4-17
	System - Adjustment.....	4-18
	System - Calibration.....	4-19
	System - Interfaces.....	4-19
	System - Accessory.....	4-20
	Installation - Information.....	4-21
	Installation - Diagnosis.....	4-21
	Stopping the Unit.....	4-22
	Power Down.....	4-22
	Shut Down.....	4-22
<b>Section 5</b>	<b>Accessories.....</b>	<b>5-1</b>
	Optional Lifting Platform Installation.....	5-1
	Optional Immersion Cooler Bridge Installation.....	5-2
	Optional Rack Assembly Instructions.....	5-3
	Optional Serial Communication Adapter.....	5-4
	Tubing.....	5-5
<b>Section 6</b>	<b>Preventive Maintenance.....</b>	<b>6-1</b>
	Cleaning.....	6-1
	Condenser Fins.....	6-1
	Testing the Safety Features.....	6-2
<b>Section 7</b>	<b>Troubleshooting.....</b>	<b>7-1</b>
	Error Displays.....	7-1
<b>Appendix</b>	<b>Serial Communications.....</b>	<b>A-1</b>

**Declaration of Conformity**

**Return Materials Authorization Form**

**Warranty**

# Preface

## Compliance

Products tested and found compliant with the requirements defined in the EC Directives as well as Low Voltage Directive (LVD) can be identified by the CE Mark on the rear of the unit. The Declaration of Conformity can be found in the appendix of this manual.

The conformity assessment were performed following defined procedures according to each applicable directive.

The council decision 93/465/EEC shall be authoritative concerning the modules of the various phases of the conformity assessment procedures and the rules for the affixing and use of the CE Mark), which are intended to be used in the technical harmonization directives.

To confirm compliance with the EC-Directive 2004/108/EC Electromagnetic Compatibility (EMC) our products were tested according to the EMC requirements for emission and immunity for electrical equipment for measurement, control and laboratory use.

Some electrical equipment displaying the CE Mark, e.g., monitors or analytical instruments, can be affected if their manufacture accepts interference, e.g, flickering of a monitor, as a minimum operating quality with electromagnetic compatibility. We recommend a minimum distance of approximately one meter from this type of equipment.

For any additional information, refer to the Declaration of Conformity.

Visit [www.ul.com/global/eng/pages](http://www.ul.com/global/eng/pages) to view the UL listings for Thermo Fisher Scientific. Reference file #E164214 for these products.

## WEEE

This product is required to comply with the European Union's Waste Electrical & Electronic Equipment (WEEE) Directive 2002/96/EC. It is marked with this symbol.



Thermo Fisher Scientific has contracted with one or more recycling/disposal companies in each EU Member State, dispose of or recycle this product through them. Further information on Thermo Fisher Scientific's compliance with these Directives is available at:

[www.thermo.com/WEEERoHS](http://www.thermo.com/WEEERoHS)


## After-sale Support

Thermo Fisher Scientific is committed to customer service both during and after the sale. If you have questions concerning the unit operation, or questions concerning spare parts or Service Contracts, call our Sales, Service and Customer Support phone number, see this manual's inside cover for contact information.



Sample Nameplate

There can be up to three nameplates located on the rear of the unit. Before calling, please obtain the serial number printed on the complete system nameplate located on the upper rear of the bath.



## Feedback

We appreciate any feedback you can give us on this manual. Please e-mail us at [thermoscientificmanuals@thermofisher.com](mailto:thermoscientificmanuals@thermofisher.com). Be sure to include the manual part number and the revision date listed on the front cover.

## Warranty

Thermo Scientific Laboratory Temperature Control Products have a warranty against defective parts and workmanship for 36 months from date of shipment. See back page of this manual for more details.

## Unpacking

Retain all cartons and packing material until the unit is operated and found to be in good condition. If the unit shows external or internal damage contact the transportation company and file a damage claim. Under ICC regulations, this is your responsibility.



**Refrigerated units should be left in an upright position for 24 hours before starting. This will ensure the lubrication oil has drained back into the compressor. ▲**

# Section 1 Safety

## Safety Warnings

Make sure you read and understand all instructions and safety precautions listed in this manual before installing or operating your unit. If you have any questions concerning the operation of your unit or the information in this manual, please contact us. See inside cover for contact information.



**DANGER** indicates an imminently hazardous situation which, if not avoided, *will* result in death or serious injury.



**WARNING** indicates a potentially hazardous situation which, if not avoided, *could* result in death or serious injury.



**CAUTION** indicates a potentially hazardous situation which, if not avoided, may result in minor or moderate injury. It is also be used to alert against unsafe practices.



The lightning flash with arrow symbol, within an equilateral triangle, is intended to alert the user to the presence of non-insulated "dangerous voltage" within the unit's enclosure. The voltage magnitude is significant enough to constitute a risk of electrical shock.



This label indicates the presence of hot surfaces.



This label indicates read the manual.

**The unit construction provides protection against the risk of electrical shock by grounding appropriate metal parts. The protection will not function unless the power cord is connected to a properly grounded outlet. It is the user's responsibility to assure a proper ground connection is provided. ▲**

**Never place the unit in a location or atmosphere where excessive heat, moisture, or corrosive materials are present. ▲**

**Never use corrosive fluids with this unit. Use of these fluids will void the manufacturer's warranty. ▲**

**Observe all warning labels. ▲**

**Never remove warning labels. ▲**

**Refrigerated units should be left in an upright position for 24 hours at room temperature before starting. This will ensure the lubrication oil has drained back into the compressor. ▲**

Other than water, before using any fluid, or when performing maintenance where contact with the fluid is likely, refer to the manufacturer's MSDS and EC Safety Data sheet for handling precautions. ▲

Ensure the tubing you select will meet your maximum temperature and pressure requirements. ▲

Never operate damaged or leaking equipment. ▲

Never operate the unit without fluid in the reservoir. ▲

Never operate the unit or add fluid to the reservoir with panels removed. ▲

Never operate the unit with the immersion circulator removed from the bath. ▲

Do not mount the immersion circulator backwards on the bath; the line cord could contact the reservoir fluid. Ensure the electrical cords do not come in contact with any of the plumbing connections or tubing. ▲

Operate the unit using only the supplied line cords, never operate equipment with damaged cords. ▲

Ensure all communication and electrical connections are made prior to starting the unit. ▲

If the unit's power cord is used as the disconnecting device, it must be easily accessible at all times. ▲

Always turn the unit off and disconnect the supply voltage from its power source before moving the unit or before performing any service or maintenance procedures. ▲

Ensure, that no toxic gases can be generated by the fluid. Inflammable gases can build up over the fluid during usage. ▲

Transport the unit with care. Sudden jolts or drops can damage the unit's components. ▲

Ensure the fluid is at a safe temperature (20°C to 55°C) before handling or draining. ▲

Drain the unit before it is transported and/or stored in near or below freezing temperatures, see Draining in Section 3. ▲

Do not clean the unit with solvents, a soft cloth and water is normally sufficient. ▲

Refer service and repairs to a qualified technician. ▲

Performance of installation, operation, or maintenance procedures other than those described in this manual may result in a hazardous situation and will void the manufacturer's warranty. ▲



## Section 2 General Information

### Description

The Thermo Scientific ADVANCED Series of thermostats are used with refrigerated and heated baths. All thermostats can pump to an external system. All controllers have a digital display and easy-to-use touch pad, five programmable setpoint temperatures, acoustic and optical alarms, and offer adjustable high temperature protection.

### ADVANCED Thermostat Specifications

	AC 150	AC 200
<b>Temperature Range</b>	Ambient +13°C to +150°C Ambient +23°F to +302°F	Ambient +13°C to +200°C Ambient +23°F to +392°F
<b>Temperature Stability</b>	±0.01°C	±0.01°C
<b>Heater Capacity</b> 230V/115V	2000/1200 Watts	2000/1200 Watts
<b>Immersion Depth</b> mm inches	75 to 145 3.0 to 5.7	75 to 145 3.0 to 5.7
<b>Dimensions</b> (H x W x D) mm inches	372 x 165 x 199 14.6 x 6.4 x 7.8	372 x 165 x 199 14.6 x 6.4 x 7.8
<b>Net Weight</b> kg lb	4.2 9.3	4.2 9.3
<b>Pumping Capacity</b> Max flow rate lpm/gpm Max pressure (mbar/psi) Max suction Pump speed steps	20/5.3 475/6.89 330/4.85 3	20/5.3 475/6.89 330/4.85 3
<b>Electrical Requirements</b> (Voltage ±10%)	100 V/50 Hz 100 V/60 Hz or 115 V/60 Hz or 230 V/50..60 Hz	100 V/50 Hz 100 V/60 Hz or 115 V/60 Hz or 230 V/50..60 Hz
<b>Connectivity</b> Remote sensor port USB port Multi function port RS232 RS485 Ethernet/LAN Analog I/O	Yes No No Optional Optional Optional No	Yes Yes Yes Optional Optional Optional Optional


- Performance specifications established in accordance with DIN 12 876 (using water at 70°C).
- Lower temperature ranges available with supplemental cooling.
- The maximum bath wall thickness for thermostats that have a factory installed clamp is 26 mm.
- Thermo Fisher Scientific reserves the right to change specifications without notice.

## ARCTIC Refrigerated Bath/Circulator Specifications

Stainless Steel Refrigerated Baths/Circulators					
	A5B	A10B	A25B	A28F	A10
<b>Temperature Range</b>	-5 to 100°C 23 to 212°F	-10 to 100°C 14 to 212°F	-25 to 200°C -13 to 392°F	-28 to 200°C -18 to 392°F	-10 to 100°C 14 to 212°F
<b>Bath Volume</b> liters gallons	12 - 20 3.2 - 5.3	17 - 30 4.5 - 7.9	13 - 21 3.4 - 5.5	6 - 10 1.6 - 2.67	4 - 6 1.1 - 1.6)
<b>Cooling Capacity</b> watts	200	250	500	320	240
<b>Refrigerant</b>	R134a	R134a	R134a	R134a	R134a
<b>Dimensions (H x W x D)*</b> mm inches	509 x 429 x 738 20.0 x 16.9 x 29.0	509 x 429 x 913 20.0 x 16.9 x 36.0	778 x 324 x 541 30.6 x 12.7 x 21.3	558 x 514 x 426 22.0 x 20.2 x 16.8	670 x 220 x 414 26.4 x 8.7 x 16.3
<b>Net Weight</b> kg lb	40.0 88.9	44.5 97.9	42.3 93.1	35.6 78.3	27.5 60.6
<b>Electrical Requirements**</b> (Voltage ±10%)	100 V/50 Hz 100 V/60 Hz or 115 V/60 Hz or 230 V/50 Hz				

Stainless Steel Refrigerated Baths/Circulators				
	A25	A28	A24B	A40
<b>Temperature Range</b>	-25 to 200°C -13 to 392°F	-28 to 200°C -18 to 392°F	-24 to 200°C -11 to 392°F	-28 to 200°C -18 to 392°F
<b>Bath Volume</b> liters gallons	7 - 12 1.8 - 3.2	6 - 10 1.6 - 2.6	16 - 27 4.2 - 7.1	7 - 12 1.8 - 3.2
<b>Cooling Capacity</b> watts	500	320	700	800
<b>Refrigerant</b>	R134a	R134a	R404	R404
<b>Dimensions (H x W x D)*</b> mm inches	749 x 273 x 483 29.5 x 10.7 x 19.0	749 x 273 x 483 29.5 x 10.7 x 19.0	612 x 765 x 610 24.1 x 30.1 x 24.0	787 x 385 x 519 31.0 x 15.1 x 20.4
<b>Net Weight</b> kg lb	36.1 79.5	36.0 79.1	58.6 128.9	55.2 121.5
<b>Electrical Requirements**</b> (Voltage ±10%)	100 V/50 Hz 100 V/60 Hz or 115 V/60 Hz or 230 V/50 Hz			

G 50 Ultra-Low Refrigerated Circulator	
<b>Temperature Range</b>	-50° to 200°C -58°F to 392°F
<b>Bath Volume</b> liters gallons	6 - 12 1.6 - 3.2
<b>Cooling Capacity at 20°C</b>	1000 watts
<b>Dimensions (H x W x D)*</b> mm inches	850 x 415 x 550 33.5 x 16.5 x 21.5
<b>Net Weight</b> kg/lb	53/117
<b>Electrical Requirements**</b> (Voltage ±10%)	200 V/50 Hz 200 V/60 Hz or 208 V/60 Hz or 230 V/50 Hz



\*Add ~26 mm (1 inch) to D for drain fitting.

\*\*See Section 3 for additional information.

• Thermo Fisher Scientific reserves the right to change specifications without notice.

## SAHARA Bath/Circulator Specifications

Stainless Steel Baths/Circulators				
	S3	S7	S13	S15
<b>Temperature Range*</b>	Ambient +13 to 200°C Ambient +23 to 392°F	Ambient +13 to 200°C Ambient +23 to 392°F	Ambient +13 to 200°C Ambient +23 to 392°F	Ambient +13 to 200°C Ambient +23 to 392°F
<b>Bath Volume</b> liters gallons	2 - 5 0.6 - 1.3	4 - 7 1.1 - 1.8	7 - 11 1.8 - 2.9	7 - 17 1.8 - 4.5
<b>Dimensions**</b> (H x W x D) mm inches	444 x 235 x 428 17.5 x 9.2 x 16.8	494 x 235 x 428 19.4 x 19.2 x 16.8	494 x 312 x 428 19.4 x 12.6 x 16.8	494 x 381 x 457 19.4 x 15.0 x 18.0
<b>Net Weight</b> kg lb	9.8 21.5	10.6 23.4	12.3 27.0	13.7 30.1

Stainless Steel Baths/Circulators				
	S21	S30	S45	S49
<b>Temperature Range*</b>	Ambient +13 to 200°C Ambient +23 to 392°F	Ambient +13 to 200°C Ambient +23 to 392°F	Ambient +13 to 200°C Ambient +23 to 392°F	Ambient +13 to 200°C Ambient +23 to 392°F
<b>Bath Volume</b> liters gallons	7 - 17 1.8 - 4.5	14 - 24 3.7 - 6.3	30 - 41 7.9 - 10.8	29 - 51 7.7 - 13.5
<b>Dimensions**</b> (H x W x D) mm inches	447 x 381 x 628 17.6 x 15.0 x 24.7	494 x 381 x 628 19.4 x 15.0 x 24.7	594 x 381 x 628 23.4 x 15.0 x 24.7	494 x 579 x 746 19.4 x 22.8 x 29.4
<b>Net Weight</b> kg lb	14.2 31.2	16.5 36.2	20.3 44.7	24.3 53.4

\*Lower temperature ranges available with supplemental cooling.

\*\*Add ~26 mm (1 inch) to D for drain fitting.

\*\*\*See Section 3 for additional information.

- Thermo Fisher Scientific reserves the right to change specifications without notice.

<b>Transparent Acrylic Baths/Circulators</b>		
	<b>S12T</b>	<b>S19T</b>
<b>Temperature Range*</b>	Ambient +13 to 60°C Ambient +23 to 140°F	Ambient +13 to 60°C Ambient +23 to 140°F
<b>Bath Volume</b> liters gallons	8 - 12 2.1 - 3.2	12 - 19 3.2 - 5.0
<b>Dimensions**</b> (H x W x D) mm inches	396 x 360 x 370 15.6 x 14.2 x 14.6	396 x 360 x 510 15.6 x 14.2 x 20.1
<b>Net Weight</b> kg lb	7.3 16.1	8.7 19.1

<b>Polyphenylene oxide (PPO) Baths/Circulators</b>		
	<b>S14P</b>	<b>S21P</b>
<b>Temperature Range*</b>	Ambient +13 to 100°C Ambient +23 to 212°F	Ambient +13 to 100°C Ambient +23 to 212°F
<b>Bath Volume</b> liters gallons	8 - 14 2.1 - 3.7	13 - 21 3.4 - 5.5
<b>Dimensions</b> (H x W x D) mm inches	396 x 360 x 450 15.6 x 14.2 x 17.7	396 x 360 x 650 15.6 x 14.2 x 25.6
<b>Net Weight</b> kg lb	6.3 13.9	6.6 14.5

\*Lower temperature ranges available with supplemental cooling.

\*\*Add ~13 mm (1/2 inch) to D for drain fitting.

\*\*\*See Section 3 for additional information.

- Thermo Fisher Scientific reserves the right to change specifications without notice.

## Wetted Materials

### ADVANCED Thermostat

Viton  
EPDM  
Ryton  
Ultem  
Vectra  
Stainless Steel

### Stainless Steel Baths/Circulators

Stainless Steel 316  
Stainless Steel 304  
EPDM (drain fitting)  
Ryton  
Zotek-N (cover seal)

### Transparent Acrylic Baths/Circulators

Poly-acryl

### Polyphenylene oxide (PPO) Baths/Circulators

Polyphenylenoxid

## Section 3 Installation

### Ambient Conditions

Ambient Temperature Range	5°C to 40°C (41°F to 104°F)
Maximum Relative Humidity	80% at 31°C (88°F)
Operating Altitude	Sea Level to 2000 meters (6560 feet)
Overvoltage Category	II
Pollution Degree	2
Degree of Protection	IP 20


---

The unit is designed for continuous operation and for indoor use.

The thermostat normally ships with it mounted facing the reservoir. You may change the position  $\pm 90^\circ$  by removing thumb screws, no tools are required.



**Do not mount it backwards; the line cord could contact the reservoir fluid. ▲**




**Never place the unit in a location where excessive heat, moisture, inadequate ventilation, or corrosive materials are present. ▲**



**Refrigerated units should be left in an upright position for 24 hours at room temperature before starting. This will ensure the lubrication oil has drained back into the compressor. ▲**

## Ventilation

The unit can operate with 0 clearance on two exhaust sides as long as the third exhaust side has unrestricted air flow. Blocked ventilation will increase the unit's temperature, reduce its cooling capacity and, on refrigerated units, eventually lead to premature compressor failure.



Ventilation Options

## Electrical Requirements



The unit construction provides protection against the risk of electrical shock by grounding appropriate metal parts. The protection will not function unless the power cord is connected to a properly grounded outlet. It is the user's responsibility to assure a proper ground connection is provided. ▲

The unit is intended for use on a dedicated outlet. All thermostats are equipped with automatic thermally-triggered 20 Amp circuit protector.

**NOTE** If the circuit protector activates allow the unit to cool before resetting. Restart the unit. Contact us if it activates again. ▲

The circuit protection is designed to protect the controller, and is not intended as a substitute for branch circuit protection. Position the unit so it is not difficult to operate the disconnecting device.



**If the unit's power cord is used as the disconnecting device, it must be easily accessible at all times. ▲**


Refer to the bath nameplate on the rear, upper-left-hand corner of the bath for specific electrical requirements. Voltage deviations of  $\pm 10\%$  are permissible. The outlet must be rated as suitable for the total power consumption of the unit, see next page.

**NOTE** If a bath and thermostat were purchased separately, follow the electrical requirements listed on the bath nameplate. ▲


The following power options are available:

Unit	Volts <sup>1</sup> /Hertz/Phase	Amps <sup>2</sup>	Total Wattage	Plug Type
<b>A10</b>	115/60/1	12	1165	N5-15
	100/50-60/1	12	1120	N5-15
	230/50/1	12	2370	Country Specific
<b>A28</b>	115/60/1	12	1185	N5-15
	100/50-60/1	12	1135	N5-15
	230/50/1	12	2395	Country Specific
<b>A25</b>	115/60/1	12	1185	N5-15
	100/50-60/1	12	1135	N5-15
	230/50/1	12	2395	Country Specific
<b>A5B</b>	115/60/1	12	1165	N5-15
	100/50-60/1	12	1120	N5-15
	230/50/1	12	2370	Country Specific
<b>A10B</b>	115/60/1	12	1165	N5-15
	100/50-60/1	12	1120	N5-15
	230/50/1	12	2370	Country Specific
<b>A25B</b>	115/60/1	12	1185	N5-15
	100/50-60/1	12	1135	N5-15
	230/50/1	12	2395	Country Specific
<b>A28F</b>	115/60/1	12	1165	N5-15
	100/50-60/1	12	1120	N5-15
	230/50/1	12	2370	Country Specific
<b>A40</b>	115/60/1	16	1660	N5-20
	100/50-60/1	16	1525	N5-20
	230/50/1	12	2600	Country Specific
<b>A24B</b>	115/60/1	16	1660	N5-20
	100/50-60/1	16	1525	N5-20
	230/50/1	12	2600	Country Specific
<b>All Heated Baths/Circulators</b>	115/60/1	12	1300	N5-20
	100/50-60/1	12	1300	N5-20
	230/50/1	12	2135	Country Specific

- 1. Volts ± 10%
- 2. Maximum amp draw



20 Amp Outlet  
(16 Amp)




15 Amp Outlet  
(12 Amp)



**For refrigerated baths:**

**Ensure all communication and electrical connections are made prior to starting the unit. ▲**

- Install the power cord from the connector on the rear of the controller, A, to the connector on the rear of the refrigerated bath, B.
- Connect the bath's power cord, C, to a grounded power outlet.




**For refrigerated baths, never connect controller power inlet, A, to a power outlet. Never connect power outlet, B, to anything but a thermostat. ▲**





**Ensure the electrical cords do not come in contact with any of the plumbing connections or tubing. ▲**


- Install the supplied communications cable between the thermostat and the bath RJ45 connectors (similar to Ethernet).





## External Circulation

The plumbing connections for external circulation are located on the rear of the thermostat.  is the return flow from the external application.  is the outlet flow to the external application (supply side). The connections are 16 mm O.D. Remove the union nuts and plates to install the 8 mm or 12 mm hose barbs and clamps supplied with the unit.



## Remote Temperature Sensor



The **External Control** on the rear of the thermostat requires a 4-pin connector that must mate to a LEMO # ECP.1S.304.CLL. The thermostat uses a 3 wire sensor, but a 4 wire sensor can be used (pins 3 and 4 are interconnected in the control head). The pin-out is:

Pin 1 and 2 = Pt100 +      Pin 3 and 4 = Pt100 -  
 See Section 4 for instructions to enable the sensor.

## Multifunction Port (AC 200 only)



The Multifunction port on the rear of the thermostat is used in conjunction with an *optional* accessory box to select and activate additional features. See Section 4 for additional information.

## USB Port (AC 200 only)

Instructions for installing the USB driver are provided on a disc included with the accessory kit.

## Tubing Requirements



**Ensure none of the tubing comes in contact with the power cord. ▲**

Tubing is normally used to connect the pump to an external application.

**NOTE** For maximum pressure to the external application cap the pump nozzle with the supplied fitting, it may be necessary to remove the thermostat from the bath to access the nozzle. ▲

**NOTE** The maximum allowable length of tube depends largely on the size, form and material of the external vessel. The length of tube and its diameter, combined with the circulating capacity, have a large effect on the temperature stability. Whenever possible, use a wider tube diameter and place the application as close as possible to the circulator. ▲



**Extreme operating temperatures will lead to extreme temperatures on the tube surface, this is even more critical with metal nozzles. ▲**

- the required tube material depends on the heat transfer liquid used
- tubes must not be folded or bent
- after prolonged use, tubes may become brittle or they may get very soft, check them on a regular basis and replace if necessary
- secure all tube connections using clamps

When using the internal bath only, the plumbing connections can be closed with the supplied plate and union nuts.

## Tubing

Tubing for Thermo Scientific temperature control systems is optional. Please select the proper tubing from the table shown in Section 5.



**Ensure the tubing you select will meet your maximum temperature and pressure requirements. ▲**

### Plastic and rubber tubing

If other plastic and rubber tubes are used, ensure that the tubes selected are fully suitable for the particular application, i.e., that they will not split, crack or become disengaged from their connections.

Connect the tubing using the supplied tube fittings for 8 or 12 mm i.d. They are attached to the plumbing connections with a supplied coupling nut.

We highly recommend using foam rubber insulation on the tubing and the fittings.

### Metal tubing

Thermo Scientific metal tubing (stainless steel insulated) offers a particularly high degree of safety and is suitable for both low and high temperatures/liquids.

The metal tubing is attached directly to the plumbing connections, gaskets are not required.



**Do not subject tubing to mechanical strain and ensure any specified bend radius is not exceeded. ▲**

Tubing is available in lengths of 0.5, 1.0 and 1.5 meters. Couplings for connecting tubes are also available.

The smallest opening inside the metal tubes is 10 mm. The metal tubing is provided with coupling nuts (M16 x 1, DIN 12 879, part 2) at either end.

## Fluids



The user is always responsible for the fluid used. Never use corrosive fluids with this unit. ▲



Never use 100% glycol. ▲



Handling and disposal of liquids other than water should be done in accordance with the fluid manufacturers specification and/or the MSDS for the fluid used. ▲



Always adjust the unit's software to the fluid used, see Section 4. ▲



When using water above 80°C closely monitor the fluid level, frequent top-offs will be required. It will also create steam. ▲



Water/glycol mixtures require top-offs with pure water, otherwise the percentage of glycol will increase resulting in high viscosity and poor performance. ▲

The fluid information is only a guide since specifications may change.

Thermo Fisher Scientific takes no responsibility for damages caused by the selection of an unsuitable bath fluid.

Unsuitable bath fluids are fluids which:

- are very highly viscous (much higher than 30 mPas at the respective working temperature)
- have corrosive characteristics or
- tend to break down at high temperatures

For fluid selection consider application requirements, operating temperature range, material compatibility, safety concerns, and environmental issues.

**5°C to 95°C — Distilled Water or Deionized Water** (up to 3 MΩ-cm)

Normal tap water leads to calcareous deposits necessitating frequent unit decalcification, see table on next page.

Calcium tends to deposit itself on the heating element. The heating capacity is reduced and service life shortened.

**-30°C to 80°C — Water with Glycol**

Below 5°C water has to be mixed with a glycol. The amount of glycol added should cover a temperature range 5°C lower than the operating temperature of the particular application. This will prevent the water/glycol from gelling (freezing) near the evaporating coil.

Excess glycol deteriorates the temperature accuracy due to its high viscosity.

**-40°C to 200°C — SIL180:**

SIL180 is suitable for covering nearly the entire range with just one liquid, especially when used with the cooling units. See next page for additional information.

Unfortunately SIL180 has a wetting tendency necessitating the occasional cleaning of the bath cover.

**other temperatures:**

Thermo Fisher Scientific offers a range of heat transfer fluids for these temperature control applications.

**SYNTH 60, SYNTH 200, SYNTH 260:**

Synthetic thermal liquid with a medium life span (several months) and little smell annoyance. See next page for additional information.

**SIL 100, SIL 180, SIL 300:**

Silicone oil with a very long life span (over 1 year) and negligible smell. See next page for additional information.

Thermo Fisher heat transfer fluids are supplied with an EC Safety Data Sheet.



**Ensure, when selecting the heat transfer fluid, that no toxic gases can be generated. Flammable gases can build up over the fluid during usage. ▲**



**Ensure the over temperature cut-off point is set lower than the fire point for the heat transfer fluid selected. ▲**



**The highest working temperature as defined by the EN 61010 (IEC 1010) must be limited to 25°C below the fire point of the bath fluid. ▲**

## **Additional Fluid Precautions**

When working with fluids other than water:

- Do not use any fluid until you have read and understood the label and the Material Safety Data Sheet (MSDS).
- Do not blend any fluids.
- Ensure any fluid residue or any other material is thoroughly removed before filling the unit with a different fluid.
- Always wear protective clothing, especially a face shield and gloves.
- Avoid spattering on any of the unit's components, always *slowly* add fluid. When adding, point the opening of a container away from yourself.
- Use fume hoods.
- Do not allow any ignition sources in the vicinity.

### Fire point

Flammable thermal liquids can ignite when a specified temperature is surpassed. The bath liquid is limited to a temperature level 25°C below the fire point as defined by the EN 61010.

### Viscosity

For optimum temperature accuracy, it is important that heat transfer liquids have a low viscosity.

#### ■ Working temperature range

This is the recommended long-term operating range. The maximum viscosity is approximately 5 mPas.


#### □ Operating temperature range

Long-term operation is recommended only under certain conditions. The viscosity may rise to a maximum of 30 mPas. The pump capacity will not match specifications.

#### ■ Heating-up range

Long-term operation is not recommended, the pump motor's excess temperature protection may switch off the pump.

Range of Application	Sil 100	Sil 180	Sil 300	Synth 60	Synth 200	Synth 260
Fire Point °C	>100	>225	>325	70	>235	275
Flash Point °C	57	170	300	59	227	260
Viscosity at 20°C (mPas)	3	11	200	2	100	140
Density at 20°C (kg/dm³)	0.89	0.93	1.08	0.76	0.86	1.03
Specific heat capacity (kJ/Kg*K)	1.67	1.51	1.56	2.10	1.96	2.00



Order Number 10 liter container	999-0202	999-0204	999-0206	999-0210	999-0226	999-0214
Order Number 5 liter container	999-0201	999-0203	999-0205	999-0209	999-0225	999-0213
Color	trans- parent, colorless	trans- parent, colorless	trans- parent, colorless	trans- parent, colorless	transpar- ent, light- brown	trans- parent, yellow
Reacts with	Silicone	Silicone	Silicone	Rubber Silicone	Copper Light metals Bronze	Copper Light metals Bronze

EC-Safety Data Sheets will be delivered together with each container of liquid.

## Water Quality and Standards

Process Fluid	Permissible (PPM)	Desirable (PPM)
<b>Microbiologicals</b>		
(algae, bacteria, fungi)	0	0
<b>Inorganic Chemicals</b>		
Calcium	<25	<0.6
Chloride	<25	<10
Copper	<1.3	<1.0
0.020 ppm if fluid in contact with aluminum		
Iron	<0.3	<0.1
Lead	<0.015	0
Magnesium	<12	<0.1
Manganese	<0.05	<0.03
Nitrates/Nitrites	<10 as N	0
Potassium	<20	<0.3
Silicate	<25	<1.0
Sodium	<20	<0.3
Sulfate	<25	<1
Hardness	<17	<0.05
Total Dissolved Solids	<50	<10
<b>Other Parameters</b>		
pH	6.5-8.5	7-8
Resistivity	0.01*	0.05-0.1*

\* MΩ-cm (compensated to 25°C)

Unfavorably high total ionized solids (TIS) can accelerate the rate of galvanic corrosion. These contaminants can function as electrolytes which increase the potential for galvanic cell corrosion and lead to localized corrosion such as pitting. Eventually, the pitting will become so extensive that refrigerant will leak into the water reservoir.

As an example, raw water in the United States averages 171 ppm (of NaCl). The recommended level for use in a water system is between 0.5 to 5.0 ppm (of NaCl).

Recommendation: Initially fill the tank with distilled or deionized water. Do not use untreated tap water as the total ionized solids level may be too high. This will reduce the electrolytic potential of the water and prevent or reduce the galvanic corrosion observed.

## Filling Requirements

Ensure the reservoir drain port on the front of the unit is *closed* and that all plumbing connections are secure. Also ensure any residue is thoroughly removed before filling the unit.



**Before using any fluid refer to the manufacturer’s MSDS and EC safety data sheets for handling precautions. ▲**

To avoid spilling, place your containers into the bath before filling.

With a low level WARNING the unit continues to run, with a FAULT the unit will shut the refrigeration, pump and heater will shut down, see Section 7. The low level warning is at approximately 3 cm (1 1/4”) below the top, the low level fault is at approximately 7.5 cm (3”).



**Avoid overfilling, oil-based fluids expand when heated. ▲**

When pumping to an external system, keep extra fluid on hand to maintain the proper level in the circulating lines and the external system.

**NOTE** Monitor the fluid level whenever heating the fluid. ▲

## Draining



**Before draining any fluid refer to the manufacturer’s MSDS and EC safety data sheets for handling precautions. ▲**



**Ensure the fluid is at a safe handling temperature, ~55°C. Wear protective clothing and gloves. ▲**

- place a suitable vessel underneath the drain. If desired, attach an 8 mm id tube on the drain.
- *slowly* turn the drain plug until flow is observed.



**Turning the drain cap more than 1 1/2 turns will result in the drain cap and fitting coming off the bath. ▲**

In this case, the drain fitting can be screwed back into the unit. Attaching the cap onto the fitting will aid in installation. If required, contact us for additional information.




Installed Drain Fitting with Cap Removed



## Section 4 Operation

### ADVANCED Thermostat

The Thermo Scientific ADVANCED Series of thermostats have a digital display and easy-to-use touch pad, five programmable setpoint temperatures, acoustic and optical alarms and some units offer adjustable high temperature protection.



This label indicates read the instruction manual before starting the unit.



Use this button to place the unit in and out of standby, see page 4-3 for more details. The blue LED illuminates when standby is enabled.



Use these navigation arrows to move through the thermostat displays and to adjust values.



Pressing this button once to make changes on the thermostat's display screen. In most cases, pressing it again is required to save the change.



Use this button to cancel any changes and to return the thermostat to its previous display. Canceling a change can only be made before the change is saved. In some cases, it is also used to save changes.



Use the adjacent dial for adjusting the High Temperature Cutout. Not all units are equipped with this feature. Details are explained in Section 6.



High Temperature Cutout reset, details are explained in Section 7.

## Setup



Refrigerated units should be left in an upright position for 24 hours before starting. This will ensure the lubrication oil has drained back into the compressor. ▲




Before starting the unit, double check all communication, electrical and plumbing connections. ▲

## Initial Start Up

Do not run the unit until fluid is added to the bath. Have extra fluid on hand. If the unit will not start refer to Section 6 Troubleshooting.

- For refrigerated units, place the circuit protector located on the rear of the bath to the **I** position.
- For all units, place the circuit protector located on the rear of the thermostat to the **I** position.
- Press . The thermostat will momentarily display: **Thermo Scientific**
- And then the Start Display will appear.



- Ensure the start symbol has a highlight box around it, if not use the arrow keys to navigate to the symbol.
- Press . The unit will start and the start symbol will turn into a stop symbol (■).



**NOTE** After start up, check all the plumbing connections for leaks. ▲

The **SP1** and **Menu** portions on the top of the display are used to view and/or change the thermostat's settings. They are explained in detail later in this Section.


**(in)** indicates the thermostat is using its internal sensor for temperature control. **(ex)** is displayed when the external sensor is selected for temperature control.

## Daily Start Up



Before starting the unit, double check all communication, electrical and plumbing connections. ▲


The blue LED on the thermostat illuminates to indicate the thermostat is in standby.


If so, press  twice. If Auto start is enabled the unit will start. If Auto start is not enabled, or if the blue LED is not illuminated, use the **Initial Start Up** procedure described on the previous page.


The Auto start feature is described later in this Section.

**NOTE** When using oil as a reservoir fluid, we recommend running the unit at 95°C for 15 minutes to remove any moisture in the fluid. ▲

## Status Display

If desired, press  to bring up the Status Display.





If desired, press  again to return to the Start Display.

## Changing the Setpoint

**NOTE** You cannot adjust the setpoint closer than 5°C to either of the fluid's system limits, see Fluids Type in this Section, or beyond the bath's temperature range, see Section 2. For example, an A10 with glycol-water cannot be set lower than -10°C or higher than +95°C. ▲


The Setpoint is the desired fluid temperature. The thermostat can store up to five setpoints, **SP1** through **SP5**. The procedure for changing the stored setpoint values is discussed later in this Section.


Use the navigation arrows and move to the **SP1** window and then press  to highlight it as shown below.





Use the up and down navigation arrows to bring up the desired setpoint and then press .

The display on the Setpoint Value Window will now indicate the corresponding setpoint's stored value.



If desired, you can change the displayed setpoint value by using the navigation arrows to highlight the Setpoint Value Window and then pressing . The right-most digit will highlight and flash.



Use the left and right arrows to move the highlight to the desired digit and then use the up and down arrows to change the value. Once all the desired changes are made, press  to save the change.


**NOTE** Using this procedure also changes the setpoint's stored value. ▲

**NOTE** The setpoint can be changed with the unit running. ▲

## Menu Displays

The thermostat uses menus to view/change the unit's settings.

**NOTE** The unit does not need to be running to view/change these settings. ▲

For all Menu displays, once  is pressed to change a display, you can press  to return to the previous screen.


1. Use the arrow buttons to highlight **Menu** and the controller brings up the Main Menu Display.

SP1	Menu
Editor	^
<b>Settings</b>	
System	
Installation	v

SP1	Menu
Editor	^
<b>Settings</b>	
System	
Installation	v

SP1	Menu
Editor	^
Settings	
<b>System</b>	
Installation	v

SP1	Menu
Editor	^
Settings	
System	
<b>Installation</b>	v

2. Use the up and down arrow to highlight the desired setting and then press  to bring up additional submenus.

Setpoints	Menu
<b>Ramp Program</b>	
Auto start	
Timer start	v
<b>Menu</b>	

See page 4-8.

Basic Settings	Menu
<b>App. Settings</b>	
Password/Reset	
<b>Menu</b>	v

See page 4-12.

Adjustment	Menu
<b>Calibration</b>	
Interfaces	
Accessory	v
<b>Menu</b>	

See page 4-16.

Servicing	Menu
<b>Information</b>	
Diagnosis	
<b>Menu</b>	v

See page 4-18.

Since the controller can only display five lines of text at a time, keep pressing the down arrow to highlight **Messages**.

SP1	^ v	Menu
Settings		^
System		
Installation		
Messages		v

Warning Messages	^
App Alarm Messages	
Safety FAULT	
	v
Menu	


See Section 7.

## Menu

The **Menu** line, at the bottom of all the submenu displays, is another way to return the thermostat back to the Main Menu Display.


1. From any submenu display, use the down arrow button to highlight **Menu**.

Setpoint	^
Ramp Program	
Auto start	
Timer start	v
Menu	

2. Press  to return to the Main Menu Display.

SP1	^ v	Menu
Editor		^
Settings		
System		
Installation		v

**NOTE**  is not operable from the Menu line. ▲



## Editor - Setpoints


**Editor Setpoints** is used to view/adjust the thermostat's five Setpoints and Real Temperature Adjustments (RTA). The display shows the temperature measured by the internal or external sensor, if this temperature does not accurately reflect the actual temperature in the reservoir an RTA is required. The RTA can be set  $\pm 10^{\circ}\text{C}$  ( $\pm 18^{\circ}\text{F}$ ).

As an example, if the thermostat temperature is stabilized and displaying  $20^{\circ}\text{C}$  but a calibrated reference thermometer reads  $20.5^{\circ}\text{C}$ , the RTA should be set to  $-0.5^{\circ}\text{C}$ . After you enter a RTA value allow unit to stabilize before verifying the temperature in the bath. **NOTE** If display accuracy is required, we recommend repeating this procedure at various setpoint temperatures and on a regular basis. ▲


**NOTE** You cannot adjust the setpoint closer than  $5^{\circ}\text{C}$  to either of the fluid's system limits, see Fluids Type in this Section, or beyond the bath's temperature range, see Section 2. For example, an A10 with glycol-water cannot be set lower than  $-10^{\circ}\text{C}$  or higher than  $+95^{\circ}\text{C}$ . ▲

1. Use the arrow buttons to highlight **Editor**.

SP1	↕	Menu
<b>Editor</b>		
Settings		
System		
Installation		

2. Press  to display the submenu.


<b>Setpoints</b>	^
Ramp Program	
Auto start	
	v
<b>Menu</b>	

3. With **Setpoints** highlighted, press  to display the list. **NOTE** Use the down arrow to display **SP5**. ▲


<b>SP1</b>	RTA	^
SP2	RTA	
SP3	RTA	
SP4	RTA	v
<b>Menu</b>		

4. The Setpoint and RTA are changed using the same procedure. With the desired setpoint highlighted press  to display the submenu.

<b>Setpoint</b>	<b>+xxx.xx</b>	^
RTA int	+xx.xx	
RTA ext	+xx.xx	
		v
<b>Menu</b>		

5. With the desired line highlighted press . A highlight will appear on the last digit.

<b>Setpoint</b>	<b>+xxx.xx</b>	^
RTA int	+xx.xx	
RTA ext	+xx.xx	
		v
<b>Menu</b>		

6. Use the left and right arrows to move the highlight to the desired digit and then use the up and down arrows to change the value. Once all the desired changes are made, press  to save the changes.



## Editor - Ramp Program (AC 200 only)


**Editor Ramp Program** is used to view/adjust the thermostat's program function.

Define your program as a series of setpoints with a known period of time interval between each. Each interval is one step of the program. Pay careful attention to the first part of your program. What conditions must exist at the beginning of your process? For example, at the starting setpoint you may wish to program an initial period of constant temperature to allow for thermal stabilization.


**NOTE** Consider the unit's limitations when designing programs. Temperature or time parameters which exceed the performance capabilities of the thermostat will result in unsatisfactory operation. If reaching the ramp setpoint temperatures is important, you will have to operate the bath between the desired setpoints and note the duration before programming the ramp.

It is possible to create a program calling for very rapid changes in temperature. Although the thermostat may not be capable of producing such changes, it may be practical to program such steps as a way to cause the fastest possible temperature change.


1. Use the arrow buttons to highlight **MENU**.




2. Use the arrow buttons to highlight **Editor**.




3. Press  to display the submenu.



4. Use the arrow buttons to highlight **Ramp Program**.



5. Press  to display:




6. With **Edit ramp** highlighted press  to display.

<b>No Of Steps</b>	<b>+XX</b>	^
<b>Variance</b>	<b>+xxx.xx</b>	
<b>Cycles</b>	<b>+xxx</b>	
		v
<b>Menu</b>		

The ramp can have up to 10 **Steps**.

The **Variance** is used to set a temperature range, the program will start when the fluid temperature is within this range. For example, if the desired **Start Temp** is 25°C and the **Variance** is set to +5°C, the program will automatically start when the bath temperature is between 20°C to 30°C.

**Cycles** sets the number of times the entire ramp program will be repeated after the last step is completed. For example, selecting 3 **Cycles** runs the entire ramp program a total of 3 times.



7. Once the **Edit ramp** portion is complete press  and then highlight **Edit step**.

<b>Edit ramp</b>	^
<b>Edit step</b>	
	v
<b>Menu</b>	

8. Use **Edit step** to enter the parameters for each step.

<b>Step #</b>	<b>+XX</b>	^
<b>Start Temp</b>	<b>+xxx.xx</b>	
<b>Stop Temp</b>	<b>+xxx.xx</b>	
<b>Duration (min)</b>	<b>+xxxxx</b>	v
<b>Menu</b>		

9. After all the desired steps are built, keep pressing  until the Start Display appears.

<b>SP1</b>	^	<b>Menu</b>
<b>+020.00</b>		
		<b>+024.29 °C</b> 
<b>Internal Temp..</b>	^	


You can enable an alarm to sound when each step and/or the program is complete, see **Settings - Basic Settings** in this section.


## Running a Ramp Program


**NOTE** The unit needs to be stopped (refrigeration and pump off) before running a ramp.








Highlighting **RAMP**, see step 2, with the unit running will cause the unit to go full cold until the ramp is started. ▲




1. Use the arrow keys to highlight the **SP** display and press .







3. Press  and the controller will display the start  and pause  program symbols. The start symbol should be highlighted .




2. Press the up arrow key until the window displays **RAMP**.





4. With the start symbol highlighted, press  to start the program. The start symbol will change into a highlighted stop program symbol .


**NOTE** With the stop symbol  highlighted, pressing  shuts down the unit. ▲


The program does not start until the process fluid temperature is at the **Step 1 Start Temp** ± the **Variance**.



5. To pause the program at the current temperature, highlight the pause program symbol  and press 


To restart the program press  again.




6. If desired, press  to bring up the Status Display. The display will indicate the current program step and the *total* time remaining for the program. When 00:0000 is displayed the program is complete and, if enabled, the alarm will sound.



7. When the program is complete the controller will maintain the last setpoint.



When  is pressed after the ramp program is complete the refrigeration, pump and controller shut off.

**NOTE** If you do not stop/end the ramp program before going back to the setpoint (**SP**) mode, changing the setpoint has no effect. For example on a program which ends with the bath at 25°C the bath will continue to hold this setpoint. Changing from ramp mode to setpoint mode and then changing the setpoint to 20°C has no effect, the bath will continue to hold 25°C. ▲

## Editor - Auto start

**Editor Auto start** is used to enable/disable auto restart. When enabled, the unit will automatically restart after a power failure or power interruption condition.

1. Use the arrow buttons to highlight **Editor**.


SP1	▲ ▼	Menu
Editor		▲
Settings		
System		
Installation		▼

2. Press  to display the submenu.


Setpoints	▲
Ramp Program	
Auto start	
Timer start	▼
Menu	


3. Use the arrows to highlight **Auto start**.

Setpoints	▲
Ramp Program	
Auto start	
Timer start	▼
Menu	

4. Press  to display the status.

Auto Start	ON	▲
		▼
Menu		

5. Press  to toggle between **ON** and **OFF**.

With the desired setting showing press , or use the **Menu** window, to save and return to the previous display.

## Editor - Timer start

**Editor Timer start** is used to set the timer and adjust the start/stop times. When enabled, the unit will start/stop using the adjusted values. To set the current time/date see the next page.

1. Press  to toggle between **ON** and **OFF**.

2. The second **Timer** is used to set the timer **Start Time/Date** and **Stop Time/Date**.

When enabled highlight the second Timer.

Timer	ON	▲
Timer		
		▼
Menu		

Start Time/Date	▲
12:00:00	
01/01/2010	
Stop Time Date	▼
Menu	

## Settings - Basic settings

**Basic settings** is used to view/adjust the thermostat's language, format and set the time and date, enable/disable the audible alarms and configure the display options.

1. Use the navigation arrow buttons to highlight **Settings**.



SP1	▲▼	Menu
Editor		▲
<b>Settings</b>		
System		
Installation		▼

2. Press  to display the submenu.

<b>Basic settings</b>	▲
App. Settings	
Password/Reset	
	▼
Menu	


3. With **Basic Settings** highlighted, press  to display the options.


<b>Language</b>	▲
Time Date	
Audible Alarms	
Display options	▼
Menu	

4. With **Language** highlighted press  and highlight the desired language. Press  (Scroll down for additional languages.)

<input checked="" type="radio"/> <b>English</b>	▲
<input type="radio"/> Deutsch	
<input type="radio"/> Francais	
<input type="radio"/> Espanol	▼
Menu	

5. With **Audible Alarms** highlighted press  to display the screen shown below.

Highlight the desired option and press  again to toggle between **OFF** and **ON**.

With the desired setting showing press , or use the **Menu** window, to save and return to the previous display.

<b>All</b>	<b>OFF</b>	▲
RAMP Prog. End	OFF	
RAMP Step End	OFF	
Safety FAULT	OFF	▼
Menu		

6. With **Display Options** highlighted press  to display the screen shown below.

Highlight the desired setting and press  to view the options.

The Temp. Units are °C, °F or °K. Temp. Resolution is either 0.01 or 0.1.

<b>Temp. Unit</b>	▲
Temp. Resolution	
Display Contrast	
Backlight	▼
Menu	

For **Display Contrast** or **Backlight** follow the instructions that appear on the screen. **Backlight** is either on or off.

## Settings - App. Settings - Limits

**App. Settings - Limits** is used to view/adjust the thermostat's high and low application temperature limits and the high and low temperature alarm limits.

1. Use the arrow buttons to highlight **App. Settings**.

Basic settings	^
<b>App. Settings</b>	
Password/Reset	v
Menu	

2. Press  to display the submenu.

<b>Limits</b>	^
Fluids Type	
	v
Menu	

3. With **Limits** highlighted, press  to display **Range Limits**.

<b>Range Limits</b>	^
Alarm Thresholds	
	v
Menu	

4. Press  and highlight the desired limit.

<b>User Temp. Limit</b>	^
System Temp. Limit	
	v
Menu	


5. Press  and highlight the desired limit.


<b>HTemp LIMIT</b>	xxx.x	^
LTemp LIMIT	xxx.x	
		v
Menu		

6. Follow the same procedure used to change the setpoint to change the limit. **NOTE** A User Temp. Limit cannot exceed a System Temp. Limit. ▲


<b>HTemp LIMIT</b>	xxx.▣	^
		v
Menu		

**NOTE** The **System Temp. Limit** is based on the fluid used, the system performance criteria and the thermostat model. It cannot be changed, see page 4-17. ▲


7. With **Alarm Thresholds** highlighted, press  to display:





9. **Define safe state** is used to select the unit's reaction to an alarm.





11. **Define Safe Temp** is used to set the safety setpoint.





8. Press  to change the thresholds.  
**NOTE** The **Thresh** limit cannot be set higher than the corresponding **Temp** limit. ▲



10. Press . Select **Run safe temp** to have the unit continue to run at a safety setpoint or select **Stop the system** to have it shut down.



12. Press  to change the setpoint.






## Settings - App. Settings - Fluids Type

**App. Settings - Fluids Type** is used to identify the type of fluid used. The thermostat uses the fluid type to automatically set certain operating parameters.

1. Use the arrow buttons to highlight **Fluids Type**.


Limits	▲
<b>Fluids Type</b>	
	▼
Menu	

2. Press  to display the submenu.

<input checked="" type="radio"/> <b>Glycol-Water</b>	▲
<input type="radio"/> <b>SIL 100</b>	
<input type="radio"/> <b>SIL 180</b>	
<input type="radio"/> <b>Water</b>	▼
Menu	

3. Highlight the applicable fluid. **NOTE** Since the controller can display only four fluid types at a time, use the down arrow to display additional fluid options. ▲

<input checked="" type="radio"/> <b>SIL 300</b>	▲
<input type="radio"/> <b>SYNTH 60</b>	
<input type="radio"/> <b>SYNTH 200</b>	
<input type="radio"/> <b>SYNTH 260</b>	▼
Menu	

With the desired setting showing press , or use the **Menu** window, to save and return to the previous display.

**NOTE** For high viscosity fluids with a wide “Heating-up range” (e.g., SIL300 - see Section 3), the thermostat automatically slows the heating from ambient to the “Operating temperature range” to help prevent fluid degradation that is caused by rapid heating while still at high viscosities. ▲

Fluid system limits:	High °C	Low °C
	AC200/AC150	AC200/AC150
<b>Water</b>	+100/+100	+5/+5
<b>Glycol-Water</b>	+100/+100	-30/-30
<b>SIL 100</b>	+75/+75	-75/-25
<b>SIL 180</b>	+200/+150	-40/-25
<b>SIL 300</b>	+200/+150	+80/+80
<b>SYNTH 60</b>	+45/+45	-50/-25
<b>SYNTH 200</b>	+200/+150	+30/+30
<b>SYNTH 260</b>	+200/+150	+45/+45
<b>Other</b>	+200/+150	-90/-90

## Settings - Password/Reset

**Password/Reset** is used to reset the thermostat back to factory preset values. **NOTE** Password is used only by a qualified technician. ▲

## System - Adjustment

**System Adjustment** is used to view/adjust the thermostat's control type, the pump speed and the level sensor and energy saving settings.

1. Use the arrow buttons to highlight **System**.

SP1	▲ ▼	Menu
Editor		▲
Settings		
<b>System</b>		
Installation		▼

2. Press  to display the submenu.

<b>Adjustment</b>	▲
Calibration	
Interfaces	
Accessory	▼
Menu	

3. With **Adjustment** highlighted, press  to display the options.

<b>Control Type</b>	▲
Pump Adjustment	
Level Sensor Settings	
Energy Saving	▼
Menu	

4. Highlight the desired control option and press  to enable it.

<input checked="" type="radio"/> <b>Internal Control</b>	▲
<input type="radio"/> External Control	
<input type="radio"/> Steaq Control	
<input type="radio"/> Follow Up Control	▼
Menu	



The **External Control** on the rear of the thermostat requires a 4-pin connector that must mate to a LEMO # ECP.1S.304.CLL. The thermostat uses a 3 wire sensor, but a 4 wire sensor can be used (pins 3 and 4 are interconnected in the control head). The pin-out is:

Pin 1 and 2 = Pt100 +      Pin 3 and 4 = Pt100 -



Control Type	▲
<b>Pump Adjustment</b>	
Level Sensor Settings	
Energy Saving	▼
Menu	

<input type="radio"/> Stage 1 (LOW)	▲
<input checked="" type="radio"/> <b>Stage 2 (MEDIUM)</b>	
<input type="radio"/> Stage 3 (HIGH)	▼
Menu	


Control Type	▲
Pump Adjustment	
<b>Level Sensor Settings</b>	
Energy Saving	▼
Menu	

<b>Lev3: NORMAL</b>	▲
Lev2: NORMAL	
Lev1: NORMAL	▼
Menu	

Highlight the desired level to change the setting. Use **Lev3** to select and enable the high level warning. **Lev2** cannot be changed. Use **Lev1** to select and enable the low level warning. See Section 6.




The Energy Saving mode is primarily designed for applications running under a stable load. Enabling the mode saves energy by reducing the unit’s heater power and cooling requirements. This can result in substantial energy savings over the life of the unit. The default setting is **ON**.

With the desired setting showing press , or use the **Menu** window, to save and return to the previous display.


## System - Calibration


**System Calibration** is used to calibrate the temperature sensors. **NOTE** Ensure the RTAs are set to 0. ▲


1. With **Calibration** highlighted, press  to display:



2. Press  to display:



3. Select the desired sensor and press  to display:



4. Enter the calibration temperature at the desired calibration point. This temperature is equal to the **Current** temperature.

Highlight the corresponding **CAL** button.

Press .

Repeat for the other two points.

## System - Interfaces

**System Interfaces** is used to view the serial communications baud rate or turn the Namur Protocol **ON** or **OFF**. This feature is optional. For additional information refer to the Appendix.

**System - Accessory** is used in conjunction with an *optional* accessory box to select and activate additional features. The box connects to the multifunction port on the rear of the thermostat.

**Boost Heater** (pins 5 and 10) is on when setpoint is >2°C above the bath temperature. Boost heater is normally off when the bath temperature is within 2°C of setpoint. **Solenoid valve for cooling coils** (pins 4 and 14) is on when the bath temperature is >2°C above setpoint or when the heater power drops to 0%. The solenoid is off when the heater power exceeds 80%. **Auto refill** (pins 5 and 10) is on when the fluid level drops below **Lev1** and is off when the fluid level is at or above **Lev3**.

**Boost Pump** (pins 4 and 14) is on whenever the option is enabled.



1. Use the arrow buttons to highlight **System**.

SP1	↕	Menu
Editor		^
Settings		
<b>System</b>		
Installation		v

2. Press to display the submenu.

<b>Adjustment</b>	^
Calibration	
Interfaces	
<b>Accessory</b>	v
Menu	

3. Highlight **Accessory** and press to display:

<b>Selection</b>	^
Activation	
External Analog Box	
	v
Menu	

4. With **Selection** highlighted, press to display:

<b>P1: +Pump / +Cooling</b>	^
<b>P2: +Heater / AutoRefill</b>	
	v
Menu	

5. With **P1** highlighted, press to display:

<input checked="" type="radio"/> <b>Booster Pump</b>	^
<input type="radio"/> Solenoid Valve	
	v
Menu	

6. With **P2** highlighted, press to display:

<input checked="" type="radio"/> <b>Booster Heater</b>	^
<input type="radio"/> Auto Refill	
	v
Menu	

7. With **Activation** highlighted, press to display the **ON/OFF** options.

Selection	^
<b>Activation</b>	
	v
Menu	

<b>Booster Pump</b>	<b>OFF</b>	^
Solenoid Valve	OFF	
Booster Heater	OFF	
Auto Refill	OFF	v
Menu		


The optional **External Analog Box** installation/operation is explained in a separate document that ships with the box.

## Installation - Information

**Installation - Information** is used to view the thermostat's software version, unit operating hours and heater/pump operating hours.

1. Use the arrow buttons to highlight **Installation**.


SP1	^ v	Menu
Editor		^
Settings		
System		
<b>Installation</b>		v

2. Press  to display the submenu.

<b>Servicing</b>	^
Information	
Diagnosis	
	v

3. Highlight **Information** and press  to display:


<b>Device info</b>	^
Operating Hours	
	v
Menu	

4. With **Device info** highlighted, press  to display the software version.

<b>Thermo Temp</b>	^
-----	
Ver.: x.xxx	
	v
Menu	

5. With **Operating Hours** highlighted, press  to display the list of components.

<b>Power Up</b>	^
Heater	
Pump	
	v
Menu	

6. Highlight the desired component, press  to display its operating hours.



<b>Hours</b>	<b>+xxxxx</b>	^
		v
Menu		


## Installation - Diagnosis

**Installation - Diagnosis** is used by a qualified technician to troubleshoot the thermostat.


## Stopping the Unit

Ensure the stop symbol has a box around it, if not use the arrow keys to navigate to the symbol.


Press . The unit will stop and the stop symbol will turn into a start symbol ().




## Power Down

Press . The thermostat's screen will go blank and the blue LED will illuminate.

## Shut Down

Place the circuit protector on the rear of the thermostat to the  position. The blue LED will extinguish.

On refrigerated units, place the circuit protector on the rear of the bath to the  position.



Using any other means to shut the unit down can reduce the life of the compressor. ▲



Always turn the unit off and disconnect it from its supply voltage before moving the unit. ▲



The circuit protector(s) located on the rear of the component(s) is not intended to act as a disconnecting means. ▲

## Restarting

**NOTE** When quickly restarting refrigerated units, the compressor may take up to 10 minutes before it starts to operate. ▲

## Section 5 Accessories

### Lifting Platform Installation




#### Tools required:

- Torx Head screwdriver
- M15 or adjustable wrench

#### Procedure:

Always turn off the unit and disconnect the power cord from the power source before installing the platform. ▲

1. Undo the four thumbscrews securing the thermostat to the bridge and remove it.
2. Undo the four Torx head screws securing the bridge to the bath and remove the bridge.
3. Secure the stacks to the platform. **NOTE** the long end of the stack is installed into the hole on the platform as shown. ▲
4. Insert the sockets into the holes on the top of the bridge. Secure the sockets to the bridge using a M15 nut on the bottom of each socket.
5. Slide the stacks up and through the sockets on the bridge.
6. Install a male knurled nut into each socket and install a female knurled nut to the top of the stack.
7. Place the assembly in the bath and secure it to the unit using the four Torx head screws.
8. Place the thermostat on the bridge and secure it using the four thumbscrews, hand tight.
9. Place the lifting platform to the desired position and lock it by using the male knurled nuts.



## Immersion Cooler Bridge Installation




### Tools required:

- Phillips Head screwdriver

### Procedure:

Always turn off the unit and disconnect the power cord from the power source before installing the bridge. ▲

1. Undo the four thumbscrews securing the thermostat to the top panel and remove the thermostat.
2. Undo the four Phillips Head screws securing the top panel to the bath and remove it.
3. Turn the old panel over and note the placement of its three gaskets. Using the old panel as a template, install the three supplied gaskets in the same position on the new panel. **NOTE** Place the panels on a soft clean cloth, their stainless steel surfaces are susceptible to scratching. ▲
4. Place the immersion cooler bridge on the bath and secure it to the unit using the four Phillips Head screws.
5. Place the thermostat on the top panel and secure it using the four thumbscrews, hand tight.
6. Remove the two screws securing the "dummy" panel to the immersion cooler bridge.
7. Insert the immersion circulator head through the hole.
8. Secure the head to the top panel using the two supplied panels.





## Rack Assembly Instructions


### Tools required:

- Phillips Head screwdriver

### Procedure:

**Note** all four support parts are identical, the lower-support is the upper-support rotated 180°.

1. If required, align the top and bottom rack supports to the desired height. Install the rack supports to the base rack using the supplied screws.
2. Install any additional racks, supplied separately, at the desired height.
3. Install optional plastic inserts into the holes as needed.



Tank Depth	Dimension A
150-160	114mm
200-205	164MM
233	188MM


### Optional Stainless Steel Insert Racks:

**A5B, A10B, A24B, S49, S19T, S14P, S21P (283 x 145 mm)**

**A25B, A410B, S21, S30 (160 x 145 mm)**

**S13, S12T (160 x 100 mm)**

- 10 mm test tube holes
- 16 mm test tube holes
- 25 mm test tube holes
- No holes



# Serial Communications Adapter




Tools required:

- None

Procedure:

Turn off the unit before installing the adapter. ▲




Adapter



Pin #	Function
1	No connection
2	TX
3	RX
4	No connection
5	GND = Signal ground
6 - 9	No connection

Pin #	Function
1-7	No connection
8	T+
9	T-

TX = Transmitted data from thermostat  
 RX = Received data to thermostat.



1. If the unit already has a communication cable installed, remove the cable from the rear of the thermostat and plug that cable into PORT 2 on the adapter.
2. Plug the supplied cable into PORT 1 on the adapter and the other end into the thermostat. Regardless of the configuration, the supplied cable *always* goes from the thermostat to PORT 1.
3. Plug the supplied serial communications cable into the communication port on the adapter and then the other end into your computer.
4. If desired, use the supplied Velcro<sup>®</sup> tape to attach the adapter to a convenient location on the unit.

## Tubing

Description	Order-No.
Insulated metal tubes made from stainless steel with M 16 x 1 unions on both ends. -90 to +105 °C temperature range	
100 cm (39") long	333-0578
150 cm (59") long	333-0579
coupling	001-2560
Insulated metal tubing made from stainless steel with M 16 x 1 unions on both ends. -50 to +300 °C temperature range	
50 cm (20") long	333-0292
100 cm (39") long	333-0293
150 cm (59") long	333-0294
tube coupling	001-2560
PVC tubing (water only)	
8 mm i.d. (available per meter)	082-0745
12 mm i.d. (available per meter)	082-0304
Viton tubing -60 to +200 °C temperature range	
8 mm i.d. (available per meter)	082-1214
12 mm i.d. (available per meter)	082-1215
Silicone tubing -30 to +220 °C temperature range (not to be used with any silicone oil, i.e., SIL or Synth 60)	
8 mm i.d. (available per meter)	082-0663
12 mm i.d. (available per meter)	082-0664
Perbunan tubing -40 to +100 °C temperature range	
8 mm i.d. (available per meter)	082-0172
12 mm i.d. (available per meter)	082-0173
Foam rubber insulation for PVC, Viton, Silicone and Perbunan tubes	
8 mm i.d. (available per meter)	806-0373
12 mm i.d. (available per meter)	806-0374
Fittings for plastic tubing	
8 mm i.d.	001-1209
12 mm i.d.	001-1210
Coupling nut	001-0797



## Section 6 Preventive Maintenance

Disconnect the power cord prior to performing any maintenance. ▲



Handle the unit with care. Sudden jolts or drops can damage the unit's components. ▲

### Cleaning

After time, the unit's stainless steel surfaces may show spots and become tarnished. Normal stainless steel cleaners can be used.

Clean the bath vessel and built-in components at least every time the bath liquid is changed. Use water and a soft cloth.



Do not use scouring powder. ▲

The inside of the bath must be kept clean in order to ensure a long service life. Substances containing acidic or alkaline substances and metal shavings should be removed quickly as they could harm the surfaces causing corrosion. If corrosion (e.g., small rust marks) should occur in spite of this, cleaning with stainless steel caustic agents has proved to be suitable. These substances should be applied according to the manufacturer's recommendations.



For cleaning the baths do not use any substances which contain solvents. ▲

### Condenser Fins

In order to maintain the cooling capacity of the unit, clean the fins two to four times per year, depending on the operating environment.

Switch off the unit and unplug the power cord.

For ARCTIC A40

- 1 Remove the condenser panel.
- 2 Clean fins with brush or similar tool.
- 3 Replace the panel.

For all other refrigerated baths:  
Clean the fins with compressed air.

For extreme soiling a qualified technician will need to remove the cooling compressor casing.

## Testing the Safety Features

The safety features for high temperature protection and low liquid level protection must be checked at regular intervals. The frequency depends on the unit's designated application and the heat transfer fluid used.



### High temperature protection

Set a cut-off temperature that is lower than the desired setpoint temperature.

Switch on the circulator and ensure the unit shuts down at the set cut-off temperature.

If the unit does not shut down have the unit checked by a qualified technician.



**Acrylic unit's maximum high temperature setting should be limited to the temperature indicated on the label on the front of the bath, 65°C or 80°C. ▲**

### Low liquid level protection

With the unit on, slowly drain the bath fluid (use a drainage tap if necessary) and ensure the unit shuts down.

If not, have the unit checked by qualified a technician.

# Section 7 Troubleshooting

## Error Displays

The thermostat can display three types of error messages.

Error messages are cleared by pressing the enter key. Once the cause of the error message is identified and corrected, to restart the unit - on units with a reset key, press the reset key full down - and then for all units press the enter key again. If **Auto start** is enabled the unit will restart, if disabled use the Start Up procedure.

### Safety Fault Displays

**NOTE** The heating element, pump and, if applicable, refrigeration shut down with a fault. ▲

A safety fault also sounds the alarm, if enabled.

#### Low Level Safety FAULT

please remove reason  
and press ENTER  
to clear message

- there is not enough liquid in the bath
- check for leaks, top off if necessary,
- fluid has evaporated, top off if necessary
- contact our Sales, Service and Customer Support

#### Temp. Safety FAULT

please remove reason  
and press ENTER  
to clear message  
Tint = xxx.xx

- high temperature protection limit exceeded
- check limit setting
- check fluid selection
- contact our Sales, Service and Customer Support

Indicates internal temperature at time of fault.



#### Motor FAULT

please remove reason  
and press ENTER  
to clear message

- it can take over 10 minutes for the motor temperature to get low enough before the unit can be restarted.
- if it quickly switches off again, contact our Sales, Service and Customer Support

#### Refrigeration FAULT

please remove reason  
and press ENTER  
to clear message

- the refrigeration may need servicing, contact our Sales, Service and Customer Support

## Warning Displays

**NOTE** The heating element and pump continue to run with a warning. ▲ A warning also sounds the alarm, if enabled.

<b>Low Level Warn.</b>
<b>please remove reason and press ENTER to clear message</b>

<b>High Level Warn.</b>
<b>please remove reason and press ENTER to clear message</b>

- check reservoir fluid level, see Section 3
- if reservoir level normal, contact our Sales, Service and Customer Support

## Other Errors

<b>A/D Converter FAULT-</b>
<b>please remove reason and press ENTER to clear message</b>

- Analog/Digital converter error, contact our Sales, Service and Customer Support

<b>Bath Communication Error</b>
<b>please remove reason and press ENTER to clear message</b>

- check connections on the back of the unit
- **NOTE** before connecting ensure the power is off. ▲
- contact our Sales, Service and Customer Support

<b>Bath not found</b>
<b>Non-refrigerated?</b>
<b>No</b>
<b>Yes</b>

- for refrigerated baths, check all communication connections
- **NOTE** before connecting ensure the power is off. ▲
- for non-refrigerated baths, select Yes
- contact our Sales, Service and Customer Support

<b>Temp. Sensor int. SHORT!</b>
<b>please remove reason and press ENTER to clear message</b>

<b>Temp. Sensor int. OPEN!</b>
<b>please remove reason and press ENTER to clear message</b>

- shorted or open internal temperature sensor
- contact our Sales, Service and Customer Support



# Appendix **Serial Communications**

## 1. Terminal settings

- Baud rate: 19200
- Data bits: 8
- Stop bits: 1
- Parity: none
- Flow control: none
- Local echo: on (the thermostat does not send a command echo)

## 2. General syntax

- Read command:
  - o Command: R<SP><parameter Id><CR><LF>
  - o Return on success: <parameter Id><value><CR><LF>
  - o Return on error: Fxxx<CR><LF>
  
- Write command:
  - o Command: W<SP><parameter Id><SP><value><CR><LF>
  - o Return on success: \$<CR><LF>
  - o Return on error: Fxxx<CR><LF>
  
- Remarks:
  - o Parameter Id consists of 2 characters
  - o Decimal point is a point “.”
  - o <SP> ↔ Space character
  - o <CR> ↔ Carriage return character
  - o <LF> ↔ Line feed character
  - o Fxxx ↔ Failure code F000 to F255
  
- Examples:
  - o W HA 0 85.5 Write parameter HA 0 with the value 85,5
  - o R LA 0 Read parameter LA 0
  - o LA +10.00 LA 0 has a value of 10,00

### 3. Table of user commands

Description	Syntax	Return on success	Namur notification
Read fluid temperature internal	R T1	T1 <value>	IN_PV_01
Read temperature setpoint	R SP	SP <value>	IN_SP_00
Write temperature setpoint	W SP <value>	\$	OUT_SP_00_<Val>
Start controller	W GO 1	\$	START
Stop controller	W RR -1	\$	STOP
Read version	R_VE	VE <value>	

### 3a. Optional commands

Description	Syntax	Return on success	Namur notification
Read temperature setpoint 1..5	R SP1 R SP2 R SP3 R SP4 R SP5	SP1 <value>	IN_SP_11 IN_SP_12 IN_SP_13 IN_SP_14 IN_SP_15
Write temperature setpoint	W SP1 <value> W SP2 <value> W SP3 <value> W SP4 <value> W SP5 <value>	\$	OUT_SP_11_<Val> OUT_SP_12_<Val> OUT_SP_13_<Val> OUT_SP_14_<Val> OUT_SP_15_<Val>
Read pump speed set point	R PU	PU01 <value>	IN_STATE_13
Write pump speed (0 to 100%)	W P5 <value>	\$	OUT_SP_01_<Val>
Read pump speed SP	R P5	P5 <Value>	IN_SP_01
Read upper temperature range limit	R HA 0	HA <value>	
Write upper temperature range limit	W HA 0 <value>	\$	
Read lower temperature range limit	R LA 0	LA <value>	
Write lower temperature range limit	W LA 0 <value>	\$	
Read upper temperature alarm limit	R HW 0	HW <value>	
Write upper temperature alarm limit	W HW 0 <value>	\$	
Read lower temperature alarm limit	R LW 0	LW <value>	
Write lower temperature alarm limit	W LW 0 <value>	\$	
Write internal RTA of setpoint 1	W C1 <value>	\$	
Read internal RTA of setpoint 1	R C1	C1 <value>	
Write internal RTA of setpoint 2	W C2 <value>	\$	
Read internal RTA of setpoint 2	R C2	C2 <value>	
Write internal RTA of setpoint 3	W C3 <value>	\$	
Read internal RTA of setpoint 3	R C3	C3 <value>	
Write internal RTA of setpoint 4	W C4 <value>	\$	
Read internal RTA of setpoint 4	R C4	C4 <value>	
Write internal RTA of setpoint 5	W C5 <value>	\$	
Read internal RTA of setpoint 5	R C5	C5 <value>	

# RMA (Return Materials Authorization) Formular / RMA Form

Die Annahme Ihres Gerätes/Ihrer Komponenten in unserem Hause kann nur erfolgen, wenn eine korrekt und vollständig ausgefüllte Erklärung mit einer gültigen RMA-Nr. vorliegt. Ist das nicht der Fall, kommt es leider zu Verzögerungen bzw. muss die Ware zurückgewiesen werden. Bitte nehmen Sie dazu unter support.mc.de@thermofisher.com Kontakt mit unserem Technischen Kunden Support auf.

*The acceptance of incoming equipment will only be carried out if a correctly completed declaration with a valid RMA no. has been submitted. Non-completion will cause a delay and the return of the equipment cannot be accepted. Please contact our Technical Support Center under support.mc.de@thermofisher.com.*

Diese Erklärung darf nur von autorisiertem Fachpersonal ausgefüllt und unterschrieben werden.

*This declaration can only be completed and signed by authorized and qualified staff:*

<p><b>1. Art der Geräte / Description of equipment</b></p> <p>Gerätetyp: Equipment type:..... Typ-Nr. BOM no.:..... Serien-Nr. Serial no.:..... RMA-Nr. RMA no.:.....</p>	<p><b>2. Grund der Einsendung / Reason for return</b></p> <p>..... ..... ..... .....</p>																					
<p><b>3. Gerätezustand / Equipment condition</b></p> <p><b>Waren die Geräte in Betrieb? / Has the equipment been used?</b></p> <p><input type="checkbox"/> Ja/Yes:                      <input type="checkbox"/> Nein/No:</p> <p><b>Sind die Geräte frei von gesundheitsgefährdenden Schadstoffen? / Is the equipment free of potentially harmful substances?</b></p> <p><input type="checkbox"/> Ja/Yes: <input type="checkbox"/> Nein/No:      (weiter Absatz 4/go to section 4)</p>	<p><b>4. Einsatzbedingte Kontaminierung / Method of contamination</b></p> <table style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 70%;"></th> <th style="width: 15%; text-align: center;">Ja/Yes</th> <th style="width: 15%; text-align: center;">Nein/No</th> </tr> </thead> <tbody> <tr> <td>- Toxisch/toxic</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>- Ätzend/corrosive</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>- Mikrobiologisch/ microbial.hazard*)</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>- Explosive/explosive*)</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>- Radioaktiv/radioactive*)</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> <tr> <td>- Sonstige Schadstoffe/ other harmful substances</td> <td style="text-align: center;"><input type="checkbox"/></td> <td style="text-align: center;"><input type="checkbox"/></td> </tr> </tbody> </table>		Ja/Yes	Nein/No	- Toxisch/toxic	<input type="checkbox"/>	<input type="checkbox"/>	- Ätzend/corrosive	<input type="checkbox"/>	<input type="checkbox"/>	- Mikrobiologisch/ microbial.hazard*)	<input type="checkbox"/>	<input type="checkbox"/>	- Explosive/explosive*)	<input type="checkbox"/>	<input type="checkbox"/>	- Radioaktiv/radioactive*)	<input type="checkbox"/>	<input type="checkbox"/>	- Sonstige Schadstoffe/ other harmful substances	<input type="checkbox"/>	<input type="checkbox"/>
	Ja/Yes	Nein/No																				
- Toxisch/toxic	<input type="checkbox"/>	<input type="checkbox"/>																				
- Ätzend/corrosive	<input type="checkbox"/>	<input type="checkbox"/>																				
- Mikrobiologisch/ microbial.hazard*)	<input type="checkbox"/>	<input type="checkbox"/>																				
- Explosive/explosive*)	<input type="checkbox"/>	<input type="checkbox"/>																				
- Radioaktiv/radioactive*)	<input type="checkbox"/>	<input type="checkbox"/>																				
- Sonstige Schadstoffe/ other harmful substances	<input type="checkbox"/>	<input type="checkbox"/>																				

**\*) Mikrobiologisch, explosiv und/oder radioaktiv kontaminierte Geräte und Komponenten werden nur bei Nachweis einer vorschriftsmäßigen Reinigung entgegengenommen!**

**We will not accept the return of any equipment that has been radioactively, explosively and/or microbiologically contaminated without written evidence of decontamination.**

Art der Schadstoffe oder prozessbedingte, gefährliche Reaktionsprodukte, mit denen die Geräte und Komponenten in Kontakt kamen:

*Please list all substances, gases and by-products which may have come into contact with the equipment:*

Handelsname/Tradename Produktname/product name Hersteller/manufacturer	Chem.Bezeichnung/ chemical name Chem.Formel/ chem. symbol	Gefahrenklasse/ hazard classification	Maßnahmen bei Freiwerden der Schadstoffe/ precautions associated with substance	Erste Hilfe bei Unfällen/ containment/ first aid measures
1.				
2.				
3.				

## Rechtsverbindliche Erklärung / Legally binding declaration

Hiermit versichere ich, dass alle gemachten Angaben korrekt und vollständig sind. Der Versand der kontaminierten Geräte erfolgt gemäß den gesetzlichen Bestimmungen.

*I hereby declare that the information supplied on this form is complete and accurate. The dispatch of equipment will be in accordance with the appropriate regulations covering packaging, transportation and labeling of dangerous substances.*

Firma/company name: .....

Adresse/address: .....

Telefon/phone: ..... Fax: .....

Ansprechpartner/contact person: ..... E-Mail: .....

Datum/Date: ..... Firmenstempel/company stamp: .....

Rechtsverbindliche Unterschrift/legally binding signature: .....

# Warranty

Thermo Fisher Scientific warrants for 36 months from date of shipment the Thermo Scientific ADVANCED series of Thermostats, ARCTIC refrigerated bath circulators, and SAHARA heated bath circulators according to the following terms.

Any part of the unit manufactured or supplied by Thermo Fisher Scientific and found in the reasonable judgment of Thermo Fisher to be defective in material or workmanship will be repaired at an authorized Thermo Fisher Repair Depot without charge for parts or labor. The unit, including any defective part must be returned to an authorized Thermo Fisher Repair Depot within the warranty period. The expense of returning the unit to the authorized Thermo Fisher Repair Depot for warranty service will be paid for by the buyer. Our responsibility in respect to warranty claims is limited to performing the required repairs or replacements, and no claim of breach of warranty shall be cause for cancellation or rescission of the contract of sales of any unit. With respect to units that qualify for field service repairs, Thermo Fisher Scientific's responsibility is limited to the component parts necessary for the repair and the labor that is required on site to perform the repair. Any travel labor or mileage charges are the financial responsibility of the buyer.

The buyer shall be responsible for any evaluation or warranty service call (including labor charges) if no defects are found with the Thermo Scientific product.

This warranty does not cover any unit that has been subject to misuse, neglect, or accident. This warranty does not apply to any damage to the unit that is the result of improper installation or maintenance, or to any unit that has been operated or maintained in any way contrary to the operating or maintenance instructions specified in this Instruction and Operation Manual. This warranty does not cover any unit that has been altered or modified so as to change its intended use.

In addition, this warranty does not extend to repairs made by the use of parts, accessories, or fluids which are either incompatible with the unit or adversely affect its operation, performance, or durability.

Thermo Fisher Scientific reserves the right to change or improve the design of any unit without assuming any obligation to modify any unit previously manufactured.

THE FOREGOING EXPRESS WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO WARRANTIES OR MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.

OUR OBLIGATION UNDER THIS WARRANTY IS STRICTLY AND EXCLUSIVELY LIMITED TO THE REPAIR OR REPLACEMENT OF DEFECTIVE COMPONENT PARTS AND Thermo Fisher Scientific DOES NOT ASSUME OR AUTHORIZE ANYONE TO ASSUME FOR IT ANY OTHER OBLIGATION.

Thermo Fisher Scientific ASSUMES NO RESPONSIBILITY FOR INCIDENTAL, CONSEQUENTIAL, OR OTHER DAMAGES INCLUDING, BUT NOT LIMITED TO LOSS OR DAMAGE TO PROPERTY, LOSS OF PROFITS OR REVENUE, LOSS OF THE UNIT, LOSS OF TIME, OR INCONVENIENCE.

This warranty applies to units sold in the United States. Any units sold elsewhere are warranted by the affiliated marketing company of Thermo Fisher Scientific. This warranty and all matters arising pursuant to it shall be governed by the law of the State of New Hampshire, United States. All legal actions brought in relation hereto shall be filed in the appropriate state or federal courts in New Hampshire, unless waived by Thermo Fisher Scientific.